



Design, Assembly and Testing of Infusion Monitoring System Based on Internet of Things

Sari Narulita¹(✉), Mohamad Syafaat², and Wulan Fitriani Safari²

¹ Faculty Nursing and Midwifery, Binawan University, East Jakarta, Indonesia
sari@binawan.ac.id

² Faculty Health Science and Technology, Binawan University, East Jakarta, Indonesia

Abstract. Intravenous therapy (infusion) or IV therapy is widely used in the medical field. IV therapy is the giving of fluids into the body, through a needle, into a vein to replace lost fluids or nutrients from the body. Monitoring of intravenous fluids is still done manually. Many cases of patient deaths are caused by negligence in monitoring intravenous fluids, to overcome this, an infusion fluid monitoring system is needed that facilitates nurse's work. The monitoring system designed must be adapted to the latest technology that may ease nurses work load. Objective. The purpose of this study is to design and realization of Internet of Thing-based infusion monitoring system. Methods. The research was carried out in several stages, system design, instrument assembly, subsystems testing Settings. The parameters observed in this system are the volume of infusion fluid (percentage) and the number of infusion fluid's drop in real time according setpoint. The system will provide a warning on the smartphone and also through a buzzer sound if something very crucial happens such as the infusion runs out. Designs. This monitoring system is equipped with sensors consisting of an infrared Light Emitting Diode (LED) and a Photodiode to observe the condition of the drops, a load cell sensor to determine the volume of infusion fluid (percentage), a servo motor to control the number of drops of infusion fluid. The sensor reading data is processed by Node MCU and then displayed on the smartphone via the Blynk application. There is a color indication in the application. If the LED is green, it means the condition is good. If the LED is yellow, it means the IV fluid is less than 20%, and if the LED is red, the IV fluid is less than 5% and the buzzer will be active. Results. IoT-based monitoring system can count infusion fluid drops with an average error of 0.73%, detecting the remaining volume of infusion fluid with an average error of 0.95% and the fastest information delivery time is 30 s. Conclusions. Internet of Things-based infusion monitoring system has been successfully made and works according to design.

Keywords: Infusion · Internet of Things · Monitoring System

1 Introduction

Intravenous therapy device (infusion) is equipment that is widely used in the medical field. Intravenous therapy is the administration of a certain amount of fluid into the body, through a needle, into a vein (vein) to lose fluids or nutrients from the body [1] protein, fat,

© The Author(s) 2023

A. Suraya et al. (Eds.): WCHSS 2022, AHSR 63, pp. 138–144, 2023.

https://doi.org/10.2991/978-94-6463-186-9_17

and calories, which cannot be adequately maintained orally, improve acid-base balance, improve the volume of blood components, provide a route for administration of drugs into the body, monitor central venous pressure (CVP), provide nutrition at the time of the digestive system is disturbed [2].

Factors that influence the success of intravenous therapy include close monitoring. Close monitoring makes it possible to replace intravenous fluids that have reached the minimum replacement limit or overcome disturbances in intravenous therapy so that complications can be avoided. Monitoring of intravenous fluids in hospitals is currently still being done manually. Medical personnel checks the volume of infusion fluids and the drip rate of infusions and infusions regularly. This is highly ineffective and has great potential for harm. Many cases of death are caused by medical personnel who are negligent in monitoring the patient's IV fluids. To overcome this, an infusion fluid monitoring system is needed that facilitates the work of medical personnel.

Efforts to develop technology to support infusion fluid monitoring are ongoing. In 2012 a system for detecting the condition of intravenous fluids was created in real-time. Infusion fluid condition detector infusion fluid volume, disturbance, and flow rate, but the system created is not connected to the internet [3] Infusion fluid monitoring systems with internet networks have been designed, but only observe the volume and rate of infusion fluids [4, 5].

Internet of Things is a structure in which objects are provided with an exclusive identity and the ability to transfer data over a network without requiring two-way human-to-human interaction but rather human-to-computer interaction [6] Internet of Things is a future science that humans use to optimize their work involving intelligent sensors and smart devices that work together via the internet [7]. IoT emerged as a big issue in the era of the industrial revolution 4.0, where it is expected that billions of physical things or objects will be equipped with various types of sensors connected to the internet. IoT is used to collect raw data in realtime in an efficient way. Where the raw data can be analyzed and processed into information needed by humans [8]. Internet of Things in its application can be used to identify, find, track, monitor objects automatically and in real time. The application of the Internet of Things has a great impact on human life in various fields, such as management, production operations, social management and even personal needs [9]. Monitoring system designed to be developed according to the needs of the latest technology medical personnel. Therefore, this research will design, manufacture and test an Internet-based infusion fluid monitoring system, where the equipment to be made can monitor the volume of infusion fluid, the flow rate of the infusion fluid, the number of drops and disturbances in real time. The data generated from the system can be accessed via the internet. If there is a condition where the infusion fluid volume is almost depleted, the flow rate and the number of infusion fluid drops that are not appropriate and their presence will be given a notification on the smartphone. The implementation of this monitoring system is expected to reduce problems that arise due to the negligence of officers.




Fig. 1. Research Steps

2 Method

This research is an experimental and descriptive research. Experimental research is at the design stage of the infusion monitoring system while descriptive research is at the stage of testing the effectiveness of the infusion monitoring system. The stages of the research and the expected achievements are as shown Fig. 1

3 Results

The research results that have been successes: designing, manufacturing, and testing an internet of thing based infusion fluid monitoring system, while testing the effectiveness of the internet of thing based infusion fluid monitoring system to patients in hospitals is still in process.

4 Discussions

4.1 Design of Infusion Fluid Monitoring System Based on Internet of Thing

The design of the system begins with making a system design which can be seen in Fig. 2.

An infrared sensor is used to determine the time of drip of infusion fluid (drops per minute and time per drop) and the number of drops of infusion fluid. A Loadcell sensor is used to determine the volume of intravenous fluids. Servo is used to adjust the infusion scroll so that the infusion flow rate is always under the set points (drops per minute) that have been given. Nodemcu is the main microcontroller that connects the monitor and control devices with the equipment. As long as the internet of thing-based infusion monitoring system equipment is connected to wifi, the process of monitoring and controlling the equipment can be carried out. The device used for control and monitoring can be an Android cellphone or computer. It could be both. The data displayed on the




Fig. 2. System Design

screen of the device that is used as control and monitoring is the volume of infusion fluid, input set points (TPM), on/off indicator of equipment, battery level, and the number of drops of infusion fluid. In addition, the data will be recorded in real-time and can be sent to email in CSV format. The data recorded are the date, time, volume of infusion fluid, set points used, and number of infusion drops. The main source of internet-based infusion monitoring system equipment is batteries. It is used to make it easier for the patient to perform activities, such as urinating.

4.2 Development of Internet of Thing Based Infusion Fluid Monitoring System

Making tools consists of assembling hardware (hardware) and software (software). Hardware assembly is done by assembling each component into a microcontroller (Nodemcu) as shown in Fig. 3.

Software assembly is done by programming on the arduino IDE. After the program is in accordance with the design, the program is downloaded to the microcontroller. The program snippet is like Fig. 4




Fig. 3. Hard Ware Assembly

```

int targetTetes = 60;
int pos = 90;
void kontrolTetes() {
  if (TPM > targetTetes) {
    if (pos > 0) pos -= 5;
    servo.write(pos);
  }
  if (TPM < targetTetes) {
    if (pos < 180) pos += 5;
    servo.write(pos);
  }
}

```

Fig. 4. Sample of Program

4.3 Testing of Infusion Monitoring System Based on Internet of Things

The tests carried out were testing the infusion fluid drop sensor, load cell sensor and monitoring. The test results of the infusion fluid drop sensor are shown in Fig. 5

Figure 5 is the result of sensor readings when infusion drops to pass through the sensor. The program shows the readings of the number of drops, the time per drop and drops per minute, and the time the infusion fluid drips. When this system was tested, the number of drops per minute was 58754, the time per drop was 0.3 and the number of

```

12:07:45.132 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:46.113 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:47.099 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:48.128 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:49.111 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:50.140 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:51.124 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:52.152 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:53.138 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:54.165 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:55.148 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:56.135 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:57.160 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:58.187 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9
12:07:59.168 -> jumlah = 587549waktPerTetes = 0.3, TetesPerMenit = 172.9

```

Fig. 5. Test Result from infusion drops Monitoring Based on Iot




Fig. 6. Test Results of the Load Cel Sensor

drops per minute was 172.9. The trial was carried out when the servo motor was open (the infusion hose was not clamped).

The loadcell sensor test results are shown in Fig. 6

Before the sensor is installed in the system, the sensor is tested first by comparing the value reading by the sensor and its comparison device (scales). The system test result is 8 while the scale is 8.5. This means that there is an error in the test results. The error value in the first test was 5.88%. The error value is obtained from the calculation results, namely $\{(8.5 - 8)/8.5\} \times 100\%$. After all the errors are calculated in each experiment, the average error is 1.26%.

The monitoring system that has been installed on the infusion pole is shown in Fig. 7




Fig. 7. Infusion Fluid Monitoring System

5 Conclusion

The Internet of Things-based infusion monitoring system has been successfully made and works according to design. Future efforts should be performed to implement this finding to bring more merit to the population by collaborating with related stakeholders, industries.

References

1. Darmadi, S. Infeksi Nosokomial Problematika & Pengendaliannya. Jkt. Salem-ba Med. (2008).
2. Potter, P. A., Perry, A. G. E., Hall, A. E. & Stockert, P. A. Fundamentals of nursing. (Elsevier Mosby, 2009).
3. Zainuri, A., Santoso, D. R. & Muslim, M. A. Monitoring dan Identifikasi Gangguan Infus Menggunakan Mikrokontroler AVR. J. EECCIS Electr. Elec-tron. Commun. Controls Inform. Syst. 6, 49–54 (2012).
4. Kusuma, T. & Mulis, M. T. Perancangan Sistem Monitoring Infus Berbasis Mikrokontroler Wemos D1 R2. Konf. Nas. Sist. Inf. KNSI 2018 (2018).
5. Agussalim, R., Adnan, A. & Niswar, M. Monitoring Cairan Infus Berdasarkan Indikator Kondisi dan Laju Cairan Infus Menggunakan Jaringan Wifi. Ilk. J. Ilm. 8, 145–152 (2016).
6. Burange, A. W. & Misalkar, H. D. Review of Internet of Things in development of smart cities with data management & privacy. in 2015 International Conference on Advances in Computer Engineering and Applications 189–195 (IEEE, 2015).
7. Sye Loong, K., Kumar, S. S. & Tshofenig, H. Securing the internet of things: A standardization perspective. Internet of Things Journal. IEEE 1, 265–275 (2014).
8. Wang, C. et al. Guest Editorial-Special issue on internet of things (IoT): Architecture, protocols and services. IEEE Sens. J. 13, 3505–3510 (2013).
9. Zhou, Q. & Zhang, J. Internet of Things and Geography-Review and Prospect. in 2011 International Conference on Multimedia and Signal Processing vol. 2 47–51 (IEEE, 2011).

Open Access This chapter is licensed under the terms of the Creative Commons Attribution-NonCommercial 4.0 International License (<http://creativecommons.org/licenses/by-nc/4.0/>), which permits any noncommercial use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license and indicate if changes were made.

The images or other third party material in this chapter are included in the chapter's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the chapter's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder.

